

File Revision Date:
September 29, 2025

Data Set Description:

PI: Alberto Redondas
Instrument: Brewer #157
Site(s): Izaña
Measurement Quantities: Total Ozone Column (TOC)

Contact Information:

Name: Alberto Redondas
Address: Centro de Investigacion Atmosferica de Izaña, La Marina 20, 38180 Spain
Phone:
FAX:
Email: aredondasm@aemet.es

Contact Information:

Name: Virgilio Carreño
Address: Centro de Investigacion Atmosferica de Izaña, La Marina 20, 38180 Spain
Phone:
FAX:
Email: vcarrenoc@aemet.es

DOI:

DOI:10.48801/eubrewnet.izana
<https://evdc.esa.int/doi/10.48801/eubrewnet.izana>

Data license:

CC-BY-NC-SA

Location of LO/L1 data archive

LO/L1 at Eubrewnet

Reference Articles:

1. Rimmer, J. S., Redondas, A. & Karppinen, T. EuBrewNet – A European Brewer network (COST Action ES1207), an overview. Atmospheric Chemistry and Physics 18, 10347–10353 (2018).
2. León-Luis, S. F. et al. Internal consistency of the Regional Brewer Calibration Centre for Europe triad during the period 2005–2016. Atmospheric Measurement Techniques 11, 4059–4072 (2018).
3. Redondas, A. et al. EUBREWNET RBCC-E Huelva 2015 Ozone Brewer Intercomparison. Atmospheric Chemistry and Physics 18, 9441–9455 (2018).
4. Grobner, J. Traceability for atmospheric total column ozone Short Name: ATMOZ, Project Number: ENV59 Final Publishable Report Traceability for atmospheric total column ozone (ENV59).

5. Redondas, A., Evans, R., Stuebi, R., Köhler, U. & Weber, M. Evaluation of the use of five laboratory determined ozone absorption cross sections in brewer and dobson retrieval algorithms. *Atmospheric Chemistry and Physics Discussions* 13, 22979–23021 (2013).
6. Kerr, J. B. The Brewer Spectrophotometer. in *UV Radiation in Global Climate Change* (eds. Gao, W., Slusser, J. R. & Schmoldt, D. L.) 160–191 (Springer Berlin Heidelberg, 2010).

Instrument Description:

Brewer MK-III (double monochromator)

The Brewer spectrophotometer was developed in Canada during the 1970s, and a commercial, automated version became available in the early 1980s. The Brewer spectrophotometer performs measurements of the direct spectral UV irradiances which, through a well-defined process, are used to calculate the TOC value. The Brewer spectrophotometer measures the direct spectral irradiance in six channels in the UV (303.2, 306.3, 310.1, 313.5, 316.8, and 320.14 nm), each with approximately a 0.5 nm bandwidth (resolving power $\lambda/\Delta\lambda \approx 600$), although that of the shortest wavelength varies with the Brewer model. The spectral analysis is achieved by a holographic grating in combination with a slit mask which selects the channel to be analyzed by a photomultiplier. There are three types of Brewer instruments: the Mk II and Mk IV models are single monochromators, and the Mk III model is a double monochromator, a characteristic that reduces stray light in its measurements. (6)

Algorithm Description:

Brewer Standard algorithm (Eubrewnet implementation)

For each ozone value five quasi-simultaneous observations are made. Each observation consists of 20 cycles of quasi-simultaneous measurements of intensity at five of the six UV-wavelengths by fast switching of the spectrometer exit slit mask.

The algorithm to retrieve TOC from individual observations is based on differential absorption of ozone at the measured wavelengths. A weighted double ratio, $R_6(O_3)$, of the measured intensities is calculated and compared to the similar extraterrestrial double ratio, ETC_0 , determined by the Langley-plot method (2) or by transfer from a travelling reference instrument (3). The weighting coefficients are designed to minimize the effects of aerosols and SO_2 spectral absorption. A differential absorption coefficient, α , corresponding to this ratio, is calculated from ozone absorption cross sections (5) convoluted with the slit functions at the measurement wavelengths, and is used to transform the $R_6(O_3)$ - ETC_0 difference into total ozone in the light's path (slant column). To get the vertical total ozone column, O_{30} , the value is further divided by the air mass factor μ

$$O_3 = (R_6(O_3) - ETC_0) / \mu \alpha$$

See more details in (1) and <http://rbcce.aemet.es/dokuwiki/doku.php?id=codes:ozonproduct>

Expected Precision/Accuracy of the Instrument:

In development, see (4)

Calibrations: Calibrated by IOS 1991-2005 every year (<https://www.io3.ca/Calibrations/Brewer/033>, <https://www.io3.ca/Calibrations/Brewer/157>). The #157 is the primary reference of the RBCC-E and is self calibrated (2) by Langley since 2011.

Instrument History:

For more details, see (2)

2005-05-31	"Change SL lamp"
2010-03-01	"Storm"
2010-09-24	"maintenance IOS10"
2012-09-05	"Hg replacement"
2012-10-24	"K&Z maintenance"
2013-04-30	"Hg replacement"
2014-02-12	"K&Z maintenance"
2014-03-17	"Wavelength Calibration"
2014-06-10	"SL shift"
2014-09-15	"SL Change Storm"
2014-12-18	"Hg replacement"
2015-04-15	"Heater issue"
2015-06-06	"maintenance IOS 2015"
2018-10-17	"Storm"
2019-04-15	"Power supply fail"
2019-05-24	"Power supply fix"
2019-10-01	"New Power Supply"
2019-10-14	"SL replacement"
2020-03-10	" Filter #4 replacement"
2020-05-24	"SL replaced"
2020-07-06	"SL cable replaced"
2020-11-09	"K&Z maintenance"
2021-02-10	"Power supply replaced"
2021-09-29	"Hg replacement"
2021-12-11	"SL change"
2022-02-23	"WL shift change"
2022-04-01	"Langley change"
2022-09-19	"Cal. 1000W"
2022-11-21	"Hg replacement."
2022-12-25	"Storm at IZO"
2023-02-25	"HR reduced"
2023-04-07	"Intensity of SL reduced by half"
2023-04-16	"SL repositioned"
2023-04-26	"Temp. control system replaced"
2023-07-31	"SL replaced"
2024-08-09	"SL replaced "
2024-11-25	"HG lamp replaced"

2025-02-14

"Flat quartz window"