

File Revision Date:

November 1, 2024

Data Set Description:

PI: Matt Tully  
Instrument: Dobson  
Site(s): Darwin, Bureau of Meteorology  
Measurement Quantities: Total column ozone

Contact Information:

Name: Dr Matt Tully  
Address: Bureau of Meteorology  
GPO Box 1289  
Melbourne, Victoria 3001  
Australia  
Phone: +61 (0)3 9669 4139  
FAX: +61 (0)3 9669 4736  
Email: Matt.Tully@bom.gov.au

Licence

CC-By-SA 4.0

DOI

Not at present time

Reference Articles:

"Operations handbook - ozone observations with a Dobson spectrophotometer", W.D. Komhyr, Global Ozone Research and Monitoring Project. Report 6, World Meteorological Organization, Geneva, 1980.

Instrument Description:

Dobson Ozone Spectrophotometer number 78

Algorithm Description:

Uses algorithm set out in "Operations handbook - ozone observations with a Dobson spectrophotometer", W.D. Komhyr, Global Ozone Research and Monitoring Project. Report 6, World Meteorological Organization, Geneva,1980.

Uses Bass/Paur (1992) ozone absorption coefficients.

Expected Precision/Accuracy of Instrument:

"Review of the Dobson spectrophotometer and its accuracy", Reid E. Basher, Global Ozone Research and Monitoring Project. Report 13, World Meteorological Organization, Geneva,1982.

Instrument History:

Dobson #078 referenced to Regional Standard 1989, 1994, 1998, 2002, 2011, 2017 and 2022