

File Revision Date:

August 21st, 2023

Data Set Description:

PI: Marie-Renee DeBacker (For Station)
Glen McConville (for instrument)

Instrument: Dobson Ozone Spectrophotometer

Site(s): l'Observatoire de Haute Provence (44N, 5.7E)

Measurement Quantities: Total Column Ozone

DOI: doi: 10.7289/V5H41PQ6

License: <https://creativecommons.org/publicdomain/zero/1.0/>

Contact Information:

Name: Dr. Marie-Renée DeBacker

Address: Groupe de Spectrométrie Moléculaire et Atmosphérique (GSMA),
Université de Reims, France

Phone: (+03) 26 91 87 77

Email: mr.debacker@univ-reims.fr

Contact Information:

Name: Glen McConville

Address: DOC/NOAA/OAR/ESRL GML, 325 Broadway, Boulder, Colorado

Phone: (+1) 720-282-9576

Email: Glen.McConville@noaa.gov

Reference Articles:

The instrument is described in numerous publications, the most commonly used reference is "Operations handbook - ozone observations with a Dobson spectrophotometer", W.D. Komhyr, Global Ozone Research and Monitoring Project. Report 183, World Meteorological Organization, Geneva, 2008.

Evans, R.D., Petropavlovskikh, I., McClure-Begley, A., McConville G., Quincy, D., and Miyagawa, K., The US Dobson Station network Data Record Prior to 2015, Re-evaluation of NDACC and WOUDC archived records with WinDobson Processing Software, Atmos. Chem. Phys., <https://doi.org/10.5194/acp-2017-383>, 2017.

Instrument Description:

Dobson Ozone Spectrophotometer Number 85 (automated)

Algorithm Description:

Uses algorithm described in "Operations handbook - ozone observations with a Dobson spectrophotometer", W.D. Komhyr, Global Ozone Research and Monitoring Project. Report 183, World Meteorological Organization, Geneva, 2008. www.esrl.noaa.gov/gmd/ozwv/dobson/GAW183-Dobson-WEB.pdf

Uses Bass/Paur ozone absorption coefficients, as defined in:
www.esrl.noaa.gov/gmd/ozwv/dobson/papers/coeffs.html

Expected Precision/Accuracy of Instrument:

There is a paper; "Review of the Dobson spectrophotometer and its accuracy", Reid E. Basher, Global Ozone Research and Monitoring Project. Report 13, World Meteorological Organization, Geneva, 1982, describing the precision and accuracy.

In general, the precision is considered to be from +/-1% (direct sun observations) to +/-5% (Observations on cloud zenith) for total ozone. Accuracy is part of an ongoing debate, but is considered in the 5% range.

Instrument History:

1983.01.01-9999.99.99 ;D085