

MetaData File provided: May 2026.

Latest Revision: May2026.

Data License:

Creative Commons Attribution 4.0 International (CC-BY-4.0)

Data Set Description:

PI up to March 1, 2026:

Martine De Mazière
Royal Belgian Institute for Space Aeronomy
Brussels, Belgium

PI starting March 1, 2026:

Corinne Vigouroux
Royal Belgian Institute for Space Aeronomy
Brussels, Belgium

Instrument: Fourier Transform Infrared Spectrometer (FTIR)

Site(s): Maito, Reunion Island
NDACC Station Maito
21.08 S, 55.38 E, 2.16 km above sea level

Measurement Quantities: Vertical Total Column Abundances above Maïdo (0-120 km) in units of
[molecules/cm²]
Vertical profiles of partial column abundances above Maïdo (0-120 km) in
units of [molecules/cm²]

Contact Information:

Name: Corinne Vigouroux
Address: Royal Belgian Institute for Space Aeronomy
3, avenue Circulaire
1180 Brussels
Belgium
Phone: +32-2-373-0-367
Email: corinne.vigouroux@aeronomie.be
WEB: <https://infrared.aeronomie.be/>

Reference Articles:

Wells, K.C., Millet, D. B., Payne, V. H., Vigouroux, C., Aquino, C. A. B., De Mazière, M., de Gouw, J. A., Graus, M., Kurosu, T., Warneke, C., and Wisthaler, A., Next-generation Isoprene Measurements from Space: Detecting Daily Variability at High Resolution, *Journal of Geophysical Research: Atmospheres*, 127, e2021JD036181, <https://doi.org/10.1029/2021JD036181>, 2022
Sha, M., B. Langerock, J.-F. L. Blavier, T. Blumenstock, T. Borsdorff, M. Buschmann, A. Dehn, M. De Mazière, N. M. Deutscher, D. G. Feist, O. E. García, D. W. T. Griffith, M. Grutter, J. W. Hannigan, F. Hase,

P. Heikkinen, C. Hermans, L. T. Iraci, P. Jeseck, N. Jones, R. Kivi, N. Kumpp, J. Landgraf, A. Lorente, E. Mahieu, M. V. Makarova, J. Mellqvist, J.-M. Metzger, I. Morino, T. Nagahama, J. Notholt, H. Ohyama, I. Ortega, M. Palm, C. Petri, D. F. Pollard, M. Rettinger, J. Robinson, S. Roche, C. M. Roehl, A. N. Röhling, C. Rousogonous, M. Schneider, K. Shiomi, D. Smale, W. Stremme, K. Strong, R. Sussmann, Y. Té, O. Uchino, V. A. Velazco, C. Vigouroux, M. Vrekoussis, P. Wang, T. Warneke, T. Wizenberg, D. Wunch, S. Yamanouchi, Y. Yang, and M. Zhou, Validation of methane and carbon monoxide from Sentinel-5 Precursor using TCCON and NDACC-IRWG stations, *Atmospheric Measurement Techniques*, 14, 6249–6304, <https://doi.org/10.5194/amt-14-6249-2021>, 2021

Vigouroux, C., Langerock, B., Bauer Aquino, C. A., Blumenstock, T., Cheng, Z., De Mazière, M., De Smedt, I., Grutter, M., Hannigan, J., Jones, N., Kivi, R., Loyola, D., Lutsch, E., Mahieu, E., Makarova, M., Metzger, J.-M., Morino, I., Murata, I., Nagahama, T., Notholt, J., Ortega, I., Palm, M., Pinardi, G., Röhling, A., Smale, D., Stremme, W., Strong, K., Sussmann, R., Té, Y., van Roozendael, M., Wang, P., and Winkler, H., TROPOMI–Sentinel-5 Precursor formaldehyde validation using an extensive network of ground-based Fourier-transform infrared stations, *Atmospheric Measurement Techniques*, 13, 3751–3767, <https://doi.org/10.5194/amt-13-3751-2020>, 2020

Zhou, M., Langerock, B., Vigouroux, C., Sha, M. K., Hermans, C., Metzger, J.-M., Chen, H., Ramonet, M., Kivi, R., Heikkinen, P., Smale, D., Pollard, D., TCCON and NDACC XCO measurements: difference, discussion and application, *Atmospheric Measurement Techniques*, 12, 5979–5995, <https://doi.org/10.5194/amt-12-5979-2019>, 2019

Zhou, X., Langerock, B., Wells, K. C., Millet, D. B., Vigouroux, C., Sha, M. K., Hermans, C., Metzger, J.-M., Kivi, R., Heikkinen, P., Smale, D., Pollard, D. F., Jones, N., Deutscher, N. M., Blumenstock, T., Schneider, M., Palm, M., Notholt, J., Hannigan, J. W., and De Mazière, M., An intercomparison of total column-averaged nitrous oxide between ground-based FTIR TCCON and NDACC measurements at seven sites and comparisons with the GEOS-Chem model, *Atmospheric Measurement Techniques*, 12, 1393–1408, <https://doi.org/10.5194/amt-12-1393-2019>, 2019

Vigouroux, C. et al.: NDACC harmonized formaldehyde time-series from 21 FTIR stations covering a wide range of column abundances, *Atmos. Meas. Tech.*, 11, 5049–5073, <https://doi.org/10.5194/amt-11-5049-2018>, 2018.

Zhou, M., Langerock, B., Vigouroux, C., Wang, P., Hermans, C., Stiller, G., Walker, K. A., Dutton, G., Mahieu, E., and De Mazière, M.: Ground-based FTIR retrievals of SF₆ on Reunion Island, *Atmos. Meas. Tech.*, 11, 651–662, <https://doi.org/10.5194/amt-11-651-2018>, 2018.

Zhou, M., Vigouroux, C., Langerock, B., Wang, P., Dutton, G., Hermans, C., Kumpp, N., Metzger, J.-M., Toon, G., and De Mazière, M.: CFC-11, CFC-12 and HCFC-22 ground-based remote sensing FTIR measurements at Réunion Island and comparisons with MIPAS/ENVISAT data, *Atmos. Meas. Tech.*, 9, 5621–5636, doi:10.5194/amt-9-5621-2016, 2016.

Baray, J.-L., Courcoux, Y., Keckhut, P., Portafaix, T., Tulet, P., Cammas, J.-P., Hauchecorne, A., Godin Beekmann, S., De Mazière, M., Hermans, C., Desmet, F., Sellegri, K., Colomb, A., Ramonet, M., Sciare, J., Vuillemin, C., Hoareau, C., Dionisi, D., Dufлот, V., Vérèmes, H., Porteneuve, J., Gabarrot, F., Gaudo, T., Metzger, J.-M., Payen, G., Leclair de Bellevue, J., Barthe, C., Posny, F., Ricaud, P., Abchiche, A., and Delmas, R.: Maïdo observatory: a new high-altitude station facility at Reunion Island (21° S, 55° E) for

long-term atmospheric remote sensing and in situ measurements, *Atmos. Meas. Tech.*, 6, 2865–2877, <https://doi.org/10.5194/amt-6-2865-2013>, 2013

Vigouroux, C., T. Stavrakou, C. Whaley, B. Dils, V. Duflot, C. Hermans, N. Kumps, J.-M. Metzger, F. Scolas, G. Vanhaelewyn, J.-F. Müller, D. B. A. Jones, Q. Li, and M. De Mazière, FTIR time-series of biomass burning products (HCN, C₂H₆, C₂H₂, CH₃OH, and HCOOH) at Reunion Island (21°S, 55°E) and comparisons with model data, *Atmos. Chem. Phys.*, 12, 10367-10385, doi:10.5194/acp-12-10367-2012, 2012.

Duflot, V., Dils, B., Baray, J. L., De Mazière, M., Attié, J. L., Vanhaelewyn, G., Senten, C., Vigouroux, C., Clain, G., and Delmas, R.: Analysis of the origin of the distribution of CO in the subtropical southern Indian Ocean in 2007, *J. Geophys. Res.*, Vol. 115, No. D22, D22106, <http://dx.doi.org/10.1029/2010JD013994>, 2010.

Vigouroux, C., Hendrick, F., Stavrakou, T., Dils, B., De Smedt, I., Hermans, C., Merlaud, A., Scolas, F., Senten, C., Vanhaelewyn, G., Fally, S., Carleer, M., Metzger, J.-M., Müller, J.-F., Van Roozendael, M., and De Mazière, M.: Ground-based FTIR and MAX-DOAS observations of formaldehyde at Réunion Island and comparisons with satellite and model data, *Atmos. Chem. Phys.*, 9, 9523-9544, 2009.

Senten, C., De Mazière, M., Dils, B., Hermans, C., Kruglanski, M., Neefs, E., Scolas, F., Vandaele, A.-C., Vanhaelewyn, G., Vigouroux, C., Carleer, M., Coheur, P.-F., Fally, S., Barret, B., Baray, J.-L., Delmas, R., Laveau, J., Metzger, J.-M., Mahieu, M., Boone, C., Walker, K. A., Bernath, P. F., and Strong K.: Technical Note: New ground-based FTIR measurements at Ile de La Réunion: observations, error analysis, and comparisons with independent data, *ACP*, 8, 3483-3508, 2008.

Neefs, E., De Mazière, M., Scolas, F., Hermans, C., and Hawat, T.: BARCOS, an automation and remote control system for atmospheric observations with a Bruker interferometer, *Review of Scientific Instruments*, 78, 3, 035109, doi:10.1063/1.2437144, 2007.

Instrument Description:

A Bruker IFS 125HR Fourier Transform Infra-red (FTIR) spectrometer has been installed at the Maito Observatory since March 2013, and performs continuous measurements since then. It is equipped with a meteo station and the BARCOS system (Neefs et al., 2007) that allows to record spectra automatically and through remote control from Brussels.

The Bruker 125HR is equipped with two detectors: InSb and MCT, and with KBr beamsplitter. The measurements cover the 750 - 5000cm⁻¹ spectral range, using seven different filter bands, as recommended in the NDACC Infrared Working Group (IRWG). A single spectrum is the average of at least one forward and one backward scan which can be taken in less than 3 minutes, but usually two or three forward & backward scans are averaged to increase the Signal to Noise Ratio. The instrument can measure with a maximum of 257 cm Optical Path Difference (OPD).

The instrument is part of [ACTRIS](https://hdl.handle.net/21.11148/3.0403c0d7d57e4567) with the following PID: <https://hdl.handle.net/21.11148/3.0403c0d7d57e4567>

Algorithm Description:

Vertical profiles of volume mixing ratios of trace gases are derived using the Optimal Estimation Method or Tikhonov regularization, as implemented in SFIT4 ([SFIT4:V0.9.4.4](#) and more recent versions [V1.01.21](#)). Vertical profiles of volume mixing ratios are weighted by the airmasses in each retrieval layer and integrated to give the total or partial columns in molecules/cm². We report total columns and profiles of partial columns.

The microwindows for NDACC target gases and interfering species follow the NDACC IRWG recommendations.

Spectra are recorded with different optical path differences, depending on the filter and the local measurement time.

Before archiving the data in the NDACC DHF, an optimized quality criterion has been applied using a threshold for the ratio of the spectral RMS residual (goodness of fit) and degrees-of-freedom for signal (DOFS). The threshold values depend on the target gas.

Ancillary Data:

Line compilation: The HITRAN line list with additional pseudo-line parameters is used in the forward calculation (except for HCHO). For interfering species the ATM line list (<http://mark4sun.jpl.nasa.gov/toon/linelist/linelist.html>) might be used instead (e.g. for H₂O). Linelists can be obtained from the SFIT4 DOIs (<https://doi.org/10.18758/clyg3euo>): currently, depending on the species, we are using

HITRAN

Rothman 2005, <https://doi.org/10.1016/j.jqsrt.2004.10.008>

Rothman 2013, <https://doi.org/10.1016/j.jqsrt.2013.07.002>

Gordon 2017, <https://doi.org/10.1016/j.jqsrt.2017.06.038>

Gordon 2022, <https://doi.org/10.1016/j.jqsrt.2021.107949>

ATM linelists

Toon 2016, <https://doi.org/10.1016/j.jqsrt.2016.05.021>

Toon 2022, <https://doi.org/10.14291/TCCON.GGG2020.ATM.R0>

Alternative download links <https://mark4sun.jpl.nasa.gov/toon/linelist/linelist.html> and <https://mark4sun.jpl.nasa.gov/pseudo.html>

Solar linelist

Hase 2005, <https://doi.org/10.1016/j.jqsrt.2006.02.026>

Physical models: A priori temperature and pressure profiles are derived from 6-hourly NCEP analyses to approx. 1.0 mbar altitude ([https://doi.org/10.1175/1520-0477\(1996\)077%3C0437:tnyrp%3E2.0.co;2](https://doi.org/10.1175/1520-0477(1996)077%3C0437:tnyrp%3E2.0.co;2)) and WACCM monthly means above (<https://doi.org/10.1029/2019jd030943>, also accessible via [V1.01.21](#)).

The Instrumental Line Shape (ILS) is monitored with HBr cell (cell #30) spectra on a quasi-regular basis. The HBr cell spectra are analysed with Linefit v.9.0 [Hase, Applied Optics, 1999]. For some retrieved gases the ILS is taken into account in the retrieval process (either in the forward model or as a retrieval parameter).

Expected Precision/Accuracy of retrieved profiles/columns:

The uncertainty evaluations are based on the methodology of Rodgers [1,2].

In addition to the measurement (S_m) errors calculated as described in those papers, random forward model parameter errors have been calculated as described by Rodgers [3] in which the K_b values are calculated by SFIT4 and our best estimates of the uncertainties in temperature (S_{temp}) and solar zenith angle (S_{sza}) are used. Systematic forward model errors, i.e. errors due to uncertainties in line intensity (S_{lint}) and line widths (S_{lwdth}), are calculated based on uncertainties provided in HITRAN. Interference errors, as described by Rodgers and Connor [4], have also been calculated to account for uncertainties in retrieval parameters (wavelength shift, instrument line shape, background slope and curvature, phase error) and in interfering gases simultaneously retrieved. These interference errors are included in the random uncertainty estimate. The total error (S_{total}) has been determined by adding all components in quadrature. For the most important contributions (temperature, measurement noise, interfering species, retrieval parameters, solar position, spectroscopy) this becomes:

$$S_{total} = \sqrt{S_m^2 + S_{temp}^2 + S_{int}^2 + S_{retparam}^2 + S_{sza}^2 + S_{lint}^2 + S_{lwdth}^2}$$

N.B. Smoothing error is not included in the error estimate.

The data user is referred to a careful discussion of error analysis for ground-based FTIR observations presented in:

[1] Rodgers CD. Retrieval of atmospheric temperature and composition from remote measurements of thermal radiation. *Rev Geophys* 14(4):609-624, 1976.

[2] Rodgers CD. Characterization and error analysis of profiles retrieved from remote sounding measurements. *J Geophys Res* 95, 5587-5595, 1990.

[3] Rodgers CD. *Inverse Methods for Atmospheric Sounding: Theory and Practice*. Series on Atmospheric, Oceanic and Planetary Physics, vol. 2. New Jersey: World Scientific Publishing Co Pte Ltd, 2000.

[4] Rodgers CD, Connor BJ. Intercomparison of remote sounding instruments. *J Geophys Res* 108, doi:10.1029/2002JD002299, 2003.

Instrument History:

Operational since March 2013. Upgraded in April 2023 to the M16 electronics system.