

File Revision Date:

September 27, 2024

Data Set Description:

PI: Margarita Yela Gonzalez

Instrument: ECC ozonesonde

Site(s): Belgrano II, Antarctica (77.86 S, 34.61 W)

Measurement Quantities: Pressure, Temperature, RH, Wind speed and direction, Geopotential height, Ozone partial pressure, Box temperature,

Contact Information:

Name: Margarita Yela Gonzalez/Monica Navarro Comas/Jose Antonio Adame Carnero

Address: Instituto Nacional de Tecnica Aeroespacial / Area de Investigacion e Instrumentacion Atmosferica

Phone: +34 915201220

Email: yelam@inta.es/navarrocm@inta.es/adamecj@inta.es

DOI:

Not at this time.

Data License:

CC-BY-SA-NC

Reference Articles:

Sepulveda, E., Cordero, R.R., Damiani, A. et al. Evaluation of Antarctic Ozone Profiles derived from OMPS-LP by using Balloon-borne Ozonesondes. *Sci Rep* 11, 4288 (2021).

<https://doi.org/10.1038/s41598-021-81954-6>

Huang, G. et al., 52/52: Validation of 10-year SAO OMI Ozone Profile (PROFOZ) product using ozonesonde observations, *Atmos. Meas. Tech.*, 10, 2455-2475, <https://doi.org/10.5194/amt-10-2455-2017>, 2017

Parrondo, M. C., Gil, M., Yela, M., Johnson, B. J., and Ochoa, H. A.: Antarctic ozone variability inside the polar vortex estimated from balloon measurements, *Atmos. Chem. Phys.*, 14, 217-229, [doi:10.5194/acp-14-217-2014](https://doi.org/10.5194/acp-14-217-2014).

Parrondo, M.C., M. Yela, M. Gil and H. Ochoa, Ozone variability at Belgrano II Station, Quadrennial Ozone Symposium, Toronto, August 27-31, 2012.

Parrondo, C., Yela, M., Gil, M., Ochoa H., Year to year ozone profile variability in the core of the Antarctic Vortex. Polar Ozone Session at Quadriennial Ozone Symposium, 2008.

Braathen G, Van der A, R., Anastou, A., Campos, J., Chipperfield, M., Ciattaglia, L., Deshler, T., Evans, R., Feng, W., Fioletov, V., Garc a, R., von der Gathen, P., Gelman, M., Ginzburg, M., Goutail, F., Hertzog, A., Johnson, B., Klekociuk, A., Koenig-Langlo, G., Long, C., Loyola, D., Manney, G., Marchand, M., McKenzie, R., McPeters, R., Mercer, J., Nash, E., Newman, P., Nichol, S., Ocampo, M., Oltmans, S., Pazmi o, A., Redondas, A., Richter, A., Rudolph, C., Shanklin, J., Shudo, Y., Vik, A.F., Weber, M., Yela, M., Zheng, X-D. Observations of the Antarctic Ozone Hole from 2003 to 2007. Polar Ozone Session at Quadriennial Ozone Symposium, 2008.

Von der Gathen, P., Rex, M., Deckelmann, H., Harris, N. R. P., Reimer, E., Grunow, K., Vik, A. F., Allaart, M., Andersen, S. B., Eriksen, P., Claude, H., David, C., Godin-Beekmann, S., Davies, J., Fast, H., De Backer, H., Deshler, T., Mercer, J., Dier, H., Dorokhov, V., Yushkov, V., Gerding, M., Johnson, B., Keckhut, P., Marchand, M., Klekociuk, A., Kivi, R., Kyr e, E., Kois, B., Zablocki, G Moore, D., Nakajima, H., Yamanouchi, T., Parrondo, C., Yela, M., Sato, K., Skrivankova, P., Viatte, P., Yamanouchi, T. Bipolar ozone loss rates measured by ozonesonde Match campaigns during IPY. Polar Ozone Session at Quadriennial Ozone Symposium , 2008.

Jiang, Y. B., L. Froidevaux, A. Lambert, N. J. Livesey, W. G. Read, J. W. Waters, B. Bojkov, T. Leblanc, I. S. McDermid, S. Godin-Beekmann, M. J. Filipiak, R. S. Harwood, R. A. Fuller, W. H. Daffer, B. J. Drouin, R. E. Cofield, D. T. Cuddy, R. F. Jarnot, B. W. Knosp, V. S. Perun, M. J. Schwartz, W. V. Snyder, P. C. Stek, R. P. Thurstans, P. A. Wagner, M. Allaart, S. B. Andersen, G. Bodeker, B. Calpini, H. Claude, G. Coetzee, J. Davies, H. De Backer, H. Dier, M. Fujiwara, B. Johnson, H. Kelder, N. P. Leme, G. Koenig-Langlo, E. Kyr e, G. Laneve, L. S. Fook, J. Merrill, G. Morris, M. Newchurch, S. Oltmans, M., C. Parrondo, F. Posny, F. Schmidlin, P. Skrivankova, R. Stubi, D. Tarasick, A. Thompson, V. Thouret, P. Viatte, H. VEmel, P. von Der Gathen, M. Yela, and G. Zablocki, Validation of Aura Microwave Limb Sounder Ozone by ozonesonde and lidar measurements, *J. Geophys. Res.*, 112, D24S34, doi:10.1029/2007JD008776, 2007

Schoeberl, M. R., J. R. Ziemke, B. Bojkov, N. Livesey, B. Duncan, S. Strahan, L. Froidevaux, S. Kulawik, P. K. Bhartia, S. Chandra, P. F. Levelt, J. C. Witte, A. M. Thompson, E. Cuevas, A. Redondas, D. W. Tarasick, J. Davies, G. Bodeker, G. Hansen, B. J. Johnson, S. J. Oltmans, H. VEmel, M. Allaart, H. Kelder, M. Newchurch, S. Godin-Beekmann, G. Ancellet, H. Claude, S. B. Andersen, E. Kyr e, M. Parrondo, M. Yela, G. Zablocki, D. Moore, H. Dier, P. von der Gathen, P. Viatte, R. Stubi, B. Calpini, P. Skrivankova, V. Dorokhov, H. de Backer, F. J. Schmidlin, G. Coetzee, M. Fujiwara, V. Thouret, F. Posny, G. Morris, J. Merrill, C. P. Leong, G. Koenig-Langlo, and E. Josephl. A trajectory-based estimate of the tropospheric ozone column using the residual method, *J. Geophys. Res.*, 112, D24S49, doi:10.1029/2007JD008773, 2007.

Yela, M., Alonso, I., Araujo, J., C rdoba, C., Deferrari, G., Diaz, S., Gil, M., Iglesias, J., Navarro, M., Parrondo, C., Puentedura, O., Ochoa, H., Rodr guez, S., Rodr guez, H., Contribuci n al proyecto del IPY: ORACLE-O3 (Ozone layer and UV radiation in a changing climate evaluated during IPY, n 99). VI Simposio Argentino y III Latinoamericano sobre investigaciones ant rticas, Buenos Aires, 10 - 14 de Septiembre, 2007.

Parrondo, M. C., Yela, M., Gil, M., von der Gathen, P. y Ochoa, H., Temperaturas observadas en la baja Estratosfera Ant rtica frente a las obtenidas por an lisis meteorol gicos y su implicaci n en la estimaci n de la destrucci n de ozono polar. VI Simposio Argentino y III Latinoamericano sobre investigaciones ant rticas, Buenos Aires, 10 - 14 de Septiembre, 2007.

Steck, T., T. von Clarmann, H. Fischer, B. Funke, N. Glatthor, U. Grabowski, M. H€epfner, S. Kellmann, M. Kiefer, A. Linden, M. Milz, G. P. Stiller, D. Y. Wang, M. Allaart, Th. Blumenstock, P. von der Gathen, G. Hansen, F. Hase, G. Hochschild, G. Kopp, E. Kyr€E, H. Oelhaf, U. Raffalski, A. Redondas Marrero, E. Remsberg, J. Russell III, K. Stebel, W. Steinbrecht, G. Wetzel, M. Yela and G. Zhang. Bias determination and precision validation of ozone profiles from MIPAS-Envisat retrieved with the IMK-IAA processor. *Atmos. Chem. Phys.*, 7, 3639-3662, 2007.

Parrondo, M.C., M. Yela, M. Gil, P. Von der Gateen and H. Ochoa, Mid-winter lower stratosphere temperatures in the Antarctic vortex: comparison between observations and ECMWF operational model, *Atmos. Chem. Phys.*, Vol. 7, pp 1-7, 2007.

Yela, M., Parrondo, C., Gil, M., Rodr€oguez, S., Araujo, J., Ochoa, H., Deferrari, G., D€az, S., Inusual comportamiento de la estratosfera antártica durante la primavera austral de 2002, *Procc. del VII Simposio Español de Estudios Polares, Granada, 18-20 September 2006.*

Parrondo, C., Yela, M., Gil, M., von der Gathen, P., Ochoa, H., Stratospheric temperatures over Antarctica in winter: Discrepancy between observations and models and its implication on the total ozone depletion estimation, *Procc. Del VII Simposio Español de Estudios Polares, Granada, 18-20 September 2006.*

Parrondo, M. C. ,Yela, M., Gil, M., Araujo, J., Ochoa, H.,Kristinsson, J., Variabilidad de la distribución de ozono en las regiones polares a partir de datos de ozonosondeos en Belgrano y Keflavik, *Procc. Del VII Simposio Español de Estudios Polares, Granada, 18-20 September 2006.*

Yela, M., C. Estevan , J. Iglesias, M. Gil, M. Navarro, R. de Mingo, C. Parrondo, O. Puentedura, S. Rodr€oguez, J. J. Araujo, H. Ochoa, S. D€az, G. Deferrari, Proyecto MARACA: Un proyecto para la Medidas de componentes Atmosféricos en la Región Antártica y ártica, *Procc. de la 5 Asamblea Hispano-Portuguesa de Geodesia y Geofísica, 30/1-3/2, Sevilla, 2006.*

Sugita, T., H. Nakajima, T. Yokota, H. Kanzawa, H. Gernandt, A. Herber, P. von der Gathen, G., K€Enig-Langlo, K. Sato, V. Dorokhov, V. A. Yushkov, Y. Murayama, M. Yamamori, S., Godin-Beekmann, F. Goutail, H. K. Roscoe, T. Deshler, M. Yela, P. Taalas, E. Kyr€E, S. J., Oltmans, B. J. Johnson, M. Allaart, Z. Litynska, A. Klekociuk, S. B. Andersen, G. O., Braathen, H. De Backer, C. E. Randall, R. M. Bevilacqua, G. Taha, L.W. Thomason, H. Irie, M. K. Ejiri, N. Saitoh, T. Tanaka, Y. Terao, H. Kobayashi, and Y. Sasano. Ozone profiles in the high-latitude stratosphere and lower mesosphere measured by the Improved Limb Atmospheric Spectrometer (ILAS)-II: Comparison with other satellite sensors and ozonesondes, *J. Geophys. Res.*, 110, D11S02, doi: 10.1029/2005JD006439, 2006.

Roscoe, H.K., S.R. Colwell , J.D. Shanklin , J.A. Karhu , P. Taalas , M. Gil , M. Yela, S. Rodriguez, C. Rafanelli, H. Cazeneuve, C.A. Villanueva, M. Ginsburg f, S.B. Diaz , R.L. de Zafra , G. Muscari, G. Redaelli, R. Dragani, Groundbased and balloon-borne measurements in Antarctic regions in support of APE-GAIA, *Advances of Space Research*, Vol: 36, 835- 845, 2005.

Yela, M., C. Parrondo, M. Gil, S. Rodr€oguez, J. Araujo, H. Ochoa, G. Deferrari and S. D€az, The September 2002 Antarctic vortex major warming as observed by visible spectroscopy and ozonesoundings, *International Journal of Remote Sensing*, V26, N16, 3361-3376, 2005.

Meijer, Y. J., D. P. J. Swart, M. Allaart, S. B. Andersen, G. Bodeker, I. Boyd, G. Braathen, Y. Calisesi, H. Claude, V. Dorokhov, P. von der Gathen, M. Gil, S. Godin-Beekmann, F. Goutail, G.

Hansen, A. Karpetchko, P. Keckhut, H. M. Kelder, R. Koelemeijer, B. Kois, R. M. Koopman, G. Kopp, J.-C. Lambert, T. Leblanc, I. S. McDe, Pole-to-pole validation of Envisat GOMOS ozone profiles using data from ground-based and balloon sonde measurements, *J. Geophys. Res.*, 109, D23305, doi: 10.1029/2004JD004834, 2004.

Parrondo, M.C. M. Yela, M. Gil, S. Rodríguez, J. Araujo, H. Ochoa, G. Deferrari y S. Døaz, Comportamiento anómalo del vórtice antártico en Septiembre de 2002 a partir de datos de espectroscopia en el rango visible y ozonosondeos, 4 Asamblea Hispano Portuguesa de Geodesia y Geofísica (Figueira da Foz, Portugal, 2 - 7 February).

Yela M., C. Parrondo, S. Rodríguez, M. Gil, J. Araujo, H. Ochoa, S. Diaz, G. Deferrari, E. Reimer, The 2002 Antarctic vortex splitting as observed from visible spectroscopy and ozonesoundings, EGS - AGU - EUG Joint Assembly (Niza, France 6-11 April). Poster.

Parrondo, M.C., M. Yela, M. P. Chipperfield, J. Araujo and M. Gil. Ozone loss rate in the lower stratosphere over Belgrano, Antarctica (78°S, 35°W) from 1999-2001 as obtained by ozonesondes: comparison with the SLIMCAT model, Sixth European Symposium on Stratospheric Ozone, (Göteborg, Sweden, 2-6 September).

Yela, M, S. Rodriguez, M. Gil, J. Iglesias, R. Ozu, S. Diaz, M. Chipperfield, Seasonal evolution of NO₂ and O₃ outside, at the edge and inside the Antarctic Polar Vortex: Comparison with the SLIMCAT model, SPARC 2000, General Assembly Stratospheric Processes and their Role in Climate (Mar del Plata, Argentina, 6 - 10 November).

Yela, M., S. Rodriguez, M. Gil, J. Iglesias, R. Ozu, S. Diaz, M. Chipperfield, NO₂ and O₃ in the core of the Antarctic vortex: Observations and modelling at Belgrano station, Quadrennial Ozone Symposium (Sapporo, Japan, 3 - 8 July), 1999.

Yela, M., S. Rodríguez, M. Gil, C. Parrondo and R. Ozu, Observaciones desde tierra en el interior del agujero de ozono antártico durante 1999, 2 Asamblea Hispano Portuguesa de Geodesia y Geofísica (Lagos, Portugal, 8 - 12 February), 1999.

Instrument Description:

The ozonesonde equipment at Belgrano is based on a receiver Marwin II DIGICORA 15. Two types of Electrochemical Concentration Cells have been used in Belgrano II along the years: Science Pump Corporation (SPC) and ENSCI Corporation.

The sonde was interfaced to a meteorological radiosonde from Vaisala for pressure, temperature and humidity data. TOTEX balloon (TX-1200) filled with helium have been used. A dip-oil treatment to reduce the occurrence of low burst height under very cold conditions are performed during the winter months.

Pre-flight procedures comply with GAW_201.pdf

https://library.wmo.int/doc_num.php?explnum_id=7167

Expected Precision/Accuracy of Instrument:

PTU values for RS 80 Radiosonde

Pressure:

Resolution 0.1 mb

Accuracy +/- 0.5 mb

Temperature:

Resolution 0.1 C

Accuracy +/- 0.2 C

Humidity:

Resolution 1% RH

Accuracy +/- 2% RH

Geopotential Height:

Uses Pressure and Temperature profile.

Errors due to uncertainty in these values.

Box Temperature:

Resolution 0.1 C

Accuracy +/- 0.5 C

Ozone Partial Pressure:

Resolution 0.01 mPa

Accuracy +/- 10% or less depending on altitude

The main sources of error are the pump correction at high altitudes and background current in the troposphere.

Instrument History:

Project start date: May, 1999

From May 1999 to June 2003: SPC Ozonesondes, model 6A have been used together with radiosondes RS80-15GE and interface cards RSA-11.

From June 2003 to November 2003: ENSCI Ozonesondes, model 2z, have been used together with radiosondes RS80-15GE and interface cards RSA-11.

Since December 2003: SPC Ozonesondes, model 6A.

Radiosonde RS92-SGPW and interface RSA 921 are using since March 2007 until 2019

Both types of sondes have been prepared following the standard operating procedures agreed upon at the NDSC ozonesonde PI meeting in Potsdam in July 1998 and the successive recommendations accepted for Standard Operation Procedures (SOPs) for QUOBI Arctic and Antarctic Match campaigns 2002-2003. A TSC- 1 ozonizer/test unit from Science Pump Corporation has been used for calibration and preparation of the sondes.

Ozonesondes preparation have always followed standard procedures with cathode solution concentration of 1% KI (full buffered), while for the ENSCI the 0.5% KI (half buffered) cathode solution was used.

Box temperatures were measured from the hole in the pump base, and the air flow rate were measured using the conventional soap bubble flow-meter.

Coefficients established by Komhyr [Komhyr 1986] were used for the altitude dependent pump efficiency correction in SPC sondes and Pump flow efficiency factors (PEF)for ENSCI Komhyr (1996)

The background current was considered constant with the altitude. Normalization to Brewer total ozone measurements has not been done, although when Brewer data at the station is available is used as a quality control parameter. Since 2007 the surface ozone monitoring TECO 49C installed is used to check the surface data of the ozonesonde before being launched. If the difference between both instruments is over 10 ppbv, the main ozonesonde parameters is checked as well as TECO data. If there is any doubt about the sonde performance, this sonde is replaced by another one.

Since 2007 the surface ozone monitoring TECO 49C installed is used to check the surface data of the ozonesonde before being launched. If the difference between both instruments is over 10 ppbv, the main ozonesonde parameters is checked as well as TECO data. If there is any doubt about the sonde performance, this sonde is replaced by another one.

In 2019 was installed new Marwin II, version 41