

File Revision Date:

October 3, 2019

Data Set Description:

PIs: Dr. Ruud Dirksen, Peter Oelsner

Instrument: Cryogenic Frostpoint Hygrometer (CFH)

Sites: Lindenberg, Germany 52.21 N, 14.12 E, 112 msl

Measurement quantities: pressure, temperature, relative humidity, geopotential height, frost point temperature, water vapor mixing ratio, mixing ratio uncertainty, vertical resolution, ozone mixing ratio, ozone partial pressure, gps altitude, latitude and longitude, horizontal wind speed and direction.

Lindenberg, Germany is also a GAW site. Simultaneous ozone measurements on the same payload are considered ancillary data. All ozone profiles should be accessed through WOUDC (<http://woudc.org/home.php>).

Contact Information:

Name: Dr. Ruud Dirksen

Address: DWD, Meteorological Observatory Lindenberg  
Am Observatorium 12, 15848 Tauche-Lindenberg, Germany

Phone: +49 69 8062 5820

Email: [Ruud.Dirksen@dwd.de](mailto:Ruud.Dirksen@dwd.de)

Name: Peter Oelsner

Address: DWD, Meteorological Observatory Lindenberg  
Am Observatorium 12, 15848 Tauche-Lindenberg, Germany

Email: [Peter.Oelsner@dwd.de](mailto:Peter.Oelsner@dwd.de)

Reference Articles:

Vömel, H., T. Naebert, R. Dirksen, and M. Sommer, (2016): An update on the uncertainties of water vapor measurements using Cryogenic Frostpoint Hygrometers, *Atmos. Meas. Tech.*, 9, 3755-3768, doi:10.5194/amt-9-3755-2016.

Vömel, H., D. E. David, and K. Smith (2007), Accuracy of tropospheric and stratospheric water vapor measurements by the cryogenic frost point hygrometer: Instrumental details and observations, *J. Geophys. Res.*, 112, D08305, doi:10.1029/2006JD007224.

Instrument Description:

The Cryogenic Frostpoint Hygrometer (CFH) is the first lightweight digital balloon-borne hygrometer based on the original NOAA analog Frostpoint Hygrometer. The CFH uses the chilled-mirror principle, in which a water condensate is formed on a small temperature-controlled mirror, which is exposed to ambient air flowing across the mirror. An optical detector senses the condensate by measuring the amount of light that is reflected off the mirror and a digital controller regulates the temperature of the mirror in order to maintain a constant reflectivity of the condensate covered mirror surface. To the extent that the reflectivity is constant, the condensate on the mirror is assumed to be in equilibrium

with the gas phase. The temperature of the mirror is measured using a small individually calibrated thermistor. Under the condition of equilibrium it is considered to be equal to the ambient dew point or frost point temperature, depending on whether the condensate phase is liquid or ice.

Algorithm Description:

The partial pressure of water vapor ( $e_w$ ) is calculated directly from the measured frost point temperature using the Goff-Gratch equation, which relates the saturation vapor pressure over ice or over liquid to the condensate temperature. The Goff Gratch equation corresponding to the correct phase of the condensate (liquid or ice) has to be used to calculate the partial pressure. The water vapor mixing ratio ( $H_2O$ ) in dry air is calculated from  $e_w$  using

$$H_2O \text{ (ppmv)} = e_w / (P - e_w) \times 10^6$$

where  $P$  is the measured atmospheric pressure.

Frost point temperatures are converted to relative humidity values by dividing the water vapor partial pressure by the saturation water vapor pressure ( $e_s$ ) at the measured atmospheric temperature.

$$RH = e_w / e_s \times 100\%$$

The uncertainty of RH values calculated in this way depends on the uncertainty of the frost point temperature measurements and the radiosonde measurements of temperature that determine  $e_s$ .

Expected Total Uncertainty of Instrument:

Vaisala RS80 Radiosonde Measurements of Pressure, Temperature and Relative Humidity

Pressure:

Total uncertainty +/- 1 hPa (at 100 hPa)

Total uncertainty +/- 0.1 hPa (at 10 hPa)

Air Temperature:

Total uncertainty +/- 0.3 K

Relative Humidity:

Total uncertainty +/- 5% RH

---

InterMet iMet-1-RSB Measurements of Pressure (after GPS correction), Temperature and Relative Humidity (PTU)

Pressure:

Total uncertainty +/- 2 hPa (at 1000 hPa)

Total uncertainty +/- 1 hPa (at 100 hPa)

Total uncertainty +/- 0.1 hPa (at 10 hPa)

Air Temperature:

Total uncertainty +/- 0.3 K

Relative Humidity:

Total uncertainty +/- 5 % RH

—

#### Vaisala RS41 Radiosonde Measurements of Pressure, Temperature and Relative Humidity

Pressure:

Total uncertainty +/- 1 hPa (> 100 hPa)

Total uncertainty +/- 0.3 hPa (at 100 to 10 hPa)

Total uncertainty +/- 0.04 hPa (at < 10 hPa)

Air Temperature:

Total uncertainty +/- 0.4 K

Relative Humidity:

Total uncertainty +/- 4% RH

—

Geopotential Height:

Calculated using radiosonde PTU measurements.

Frost Point Temperature:

Total uncertainty +/- 0.1 K

Water Vapor Mixing Ratio:

Total uncertainty typically +/- 2 % (1 sigma)

The total uncertainty is provided as additional column within the data.

The vertical width of the smoothing kernel for which this uncertainty applies is also provided as part of the data.

For the algorithm to estimate the water vapor uncertainty see Vömel et al., (2016)

#### Measurement History:

First observation	August 2006
Regular observations started	October 2008
Transition to InterMet iMet-1-RSB radiosonde	July 2009
Switch from monthly to bi-weekly observations	April 2010
Transition to Vaisala RS41 radiosonde	February 2016